

This is to certify that the equipment listed has been examined and tested in accordance with the requirements of:-

- a) The MMC temperature specification and test procedures, using ASTM or equivalent thermometers (certificate No's 103161001 to 103161004 traceable to National Standards).
- b) The MMC linear Specification and test procedures, using steel tape/rule (certificate No's 06031/06047 traceable to National Standards).

ACTUAL READING:	0.1 °C	23.3 °C	52.2 °C	75.6 °C	93.0 cms
UNIT UNDER TEST:	0.1 °C	23.3 °C	52.1 °C	75.6 °C	92.7 cms
DIFFERENCE:	0.0 °C	0.0 °C	-0.1 °C	0.0 °C	-0.3 cms

The equipment tested comprises of the following components, and this certificate is only valid for this equipment.

TYPE OR MODEL : D-2401-2

DESCRIPTION : 30m MMC Triple Function Gauging Unit

SERIAL NO : 20522

SIGNED :



FOR AND ON BEHALF OF MMC (EUROPE) LTD